

# Certificate of Analysis

**Product** H50V DILUENT, RFID  
**Product No** 1440103 (5 L)  
**Lot No** 24050416M3

**Manufacturing Date** 2024-05-20

**Expiration Date** 2026-05-19

Test	Specification
Appearance	Diluent should be transparent liquid without precipitation, particle or floccules
pH	6.80 – 7.00
Osmolality (mOsm/kg)	309 – 321
Conductivity (mS/cm)	18.50 - 19.50
Particles (10 <sup>5</sup> /L) (Particle volume≥2.5fL)	≤ 2.5

Actual values meet specifications.

Date: 2024-10-17



Ulrika Miefalk  
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.  
All information in this document is property of Boule Medical AB.

