


Certificate of Analysis

Product MEDONIC M51-D DILUENT (20L)
Product No 1504510
Lot No 2024072302
Manufacturing Date 2024-07-23
Expiration Date 2026-07-22

| Test | Specification |
|--|---------------|
| pH | 7,1 – 7,5 |
| Osmolality (mOsm/kg) | 290 – 310 |
| Conductivity (mS/cm) | 17,3 – 18,3 |
| Particles PLT/RBC channel (x 10 ⁹ /L) | ≤ 5 |

Actual values meet specification.

Date: 2024-10-30


Ulrika Miefalk
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

