

Boule Con-Diff

CONTROL

IVD



Control	Lot	
Tri-Level	22410-1K	2025-03-03
Low	22410-11	2025-03-03
Normal	22410-12	2025-03-03
High	22410-13	2025-03-06

Open vial stability 14 days ¹

MEDONIC		CONTROL	L
Medonic M-series/M32		LOT	22410-11
		Mean ²	± Range ³
RBC	10 ¹² /L	2.27	0.12
MCV	fL	66.2	5.0
RDW	%	15.4	4.0
Hct	%	15.0	2.5
Plt	10 ⁹ /L	86	15
MPV	fL	10.6	1.5
WBC	10 ⁹ /L	3.5	0.3
Hgb	g/dL	5.6	0.3
MCH	pg	24.7	1.5
MCHC	g/dL	37.3	3.0
Lymph#	10 ⁹ /L	1.6	0.4
Mid#	10 ⁹ /L	0.2	0.3
Gran#	10 ⁹ /L	1.7	0.8
Lymph	%	45.6	7.0
Mid	%	5.6	5.0
Gran	%	48.8	7.0

CONTROL	N	
LOT	22410-12	
	Mean ²	± Range ³
	4.14	0.18
	78.8	5.0
	14.9	4.0
	32.6	3.0
	238	30
	10.9	1.5
	8.5	0.6
	11.5	0.4
	27.8	1.5
	35.3	3.0
	4.0	0.6
	0.6	0.5
	3.9	1.0
	47.3	5.0
	7.2	5.0
	45.5	5.0

CONTROL	H	
LOT	22410-13	
	Mean ²	± Range ³
	5.11	0.22
	93.7	5.0
	13.9	4.0
	47.9	3.5
	521	60
	11.3	1.5
	20.6	1.8
	16.3	0.5
	31.9	1.5
	34.0	3.0
	12.6	1.5
	1.2	1.0
	6.8	2.5
	61.0	5.0
	5.9	5.0
	33.1	5.0

- 1
- bg Трайност на отворен флакон 14 дни (bg)
 - hr Stabilnost otvorene bočice 14 dani (hr)
 - cs Stabilita otevřené ampulky 14 dní (cs)
 - da Åben rør stabilitet 14 dage (da)
 - et Stabiilsus avatud vialis 14 päevade (et)
 - fr Stabilité en flacon ouvert 14 jour (fr)
 - de Stabilität geöffneter Flaschen 14 tage (de)
 - el Διατήρηση δείγματος μετά την αποσφράγιση 14 ημέρα (el)
 - it Stabilità della fiala aperta 14 giorni (it)
 - lv Atvērt flakonu stabilitāti 14 dienas (lv)
 - lt Stabilumas atidarius buteliuką 14 dienas (lt)
 - no Åpen rør/glass stabilitet 14 dager (no)
 - pl Trwałość po otwarciu fiołki 14 Liczba dni (pl)
 - pt Estabilidade após abertura do frasco 14 dias (pt)
 - ro Termenul de valabilitate al fiolei desfăcute 14 zile (ro)
 - ru Стабильность вскрытого флакона 14 дней (ru)
 - sr Proizvod stabilan nakon otvaranja pakovanja 14 dana (sr)
 - sk Stabilita otvorenej injekčnej liekovky 14 dní (sk)
 - sl Stabilita otvorenej injekčnej liekovky 14 dni (sl)
 - es Estabilidad de la cápsula abierta 14 días (es)
 - sv Hållbarhet för öppnad flaska 14 dagar (sv)
 - tr Açık şişe stabilitesi 14 gün (tr)

- 2
- Средно
 - Sredstvo
 - Průměr
 - Middelværdi
 - Keskmine
 - Moyenne
 - Mittelwert
 - Μέση
 - Media
 - Nozīmē
 - Vidurkis
 - Middelverdi
 - Średnia
 - Média
 - Medie
 - Среднее
 - Srednja vrednost
 - Priemer
 - Povprečno
 - Media
 - Medelvärde
 - Ortalama

- 3
- Обхват
 - Raspon
 - Rozsah
 - Område
 - Vahemik
 - Intervalle
 - Bereich
 - Αναμενόμενο εύρος α
 - Intervallo
 - Diapazons
 - Intervalas
 - Område
 - Zakres
 - Intervalo
 - Interval
 - диапазон
 - Opseg
 - Rozsah
 - Območje
 - Intervallo
 - Intervall
 - Aralığı

For Medonic M-series / M32 with barcode reader




M-Series:

1. In the Main Menu go to "Q/C Menu" "Enter Con/Cal" "Control/Calibrator Assay Value Input"
2. Scan in each blood calibrator/control column by column, sequence 1 to 9

M32:

1. In the Main Menu go to "Quality Control" "Input Assays"
2. Scan in each blood calibrator / control column by column, sequence 1 to 9

Instr.: M-series	Instr.: M-series	Instr.: M-series
Lot: 22410-11	Lot: 22410-12	Lot: 22410-13
Prod.: Con-Diff Low	Prod.: Con-Diff Norm	Prod.: Con-Diff High
ID: 2241011+	ID: 2241012+	ID: 2241013+
Exp.: 3-Mar-2025	Exp.: 3-Mar-2025	Exp.: 6-Mar-2025

1:  Boul eCBAV2	 Boul eCBAV2	 Boul eCBAV2
2:  2241011+	 2241012+	 2241013+
3:  25030300001153030000	 25030300002153030000	 25030600003153030000
4:  10004031100000000000	 10004030400000000000	 10004030200000000000
5:  22712662500560315025	 41418788501150432630	 51122937501630547935
6:  08615106152471537330	 23830109152781535330	 52160113153191534030
7:  03503016040020301708	 08506040060060503910	 20618126150121006825
8:  15440456700565048870	 14940473500725045550	 13940610500595033150
9:  11117058318013471137	 11117401944939535278	 11119341637629760383

Boule Con-Diff

CONTROL

IVD



Control	Lot	
Tri-Level	22410-1K	2025-03-03
Low	22410-11	2025-03-03
Normal	22410-12	2025-03-03
High	22410-13	2025-03-06

Open vial stability 14 days ¹

SWELAB

Swelab Alfa/Alfa Plus	CONTROL LOT	L 22410-11	
		Mean ²	± Range ³
RBC	10 ¹² /L	2.31	0.12
MCV	fL	68.2	5.0
RDW	%	14.9	4.0
Hct	%	15.8	2.5
Plt	10 ⁹ /L	86	15
MPV	fL	11.9	1.5
WBC	10 ⁹ /L	3.3	0.3
Hgb	g/dL	5.7	0.3
MCH	pg	24.7	1.5
MCHC	g/dL	36.1	3.0
Lymph#	10 ⁹ /L	1.6	0.4
Mid#	10 ⁹ /L	0.1	0.3
Gran#	10 ⁹ /L	1.6	0.8
Lymph	%	48.5	7.0
Mid	%	4.1	5.0
Gran	%	47.4	7.0

CONTROL LOT	N 22410-12	
	Mean ²	± Range ³
	4.17	0.18
	80.7	5.0
	13.7	4.0
	33.7	3.0
	244	30
	12.0	1.5
	8.5	0.6
	11.6	0.4
	27.8	1.5
	34.4	3.0
	4.3	0.6
	0.4	0.5
	3.8	1.0
	51.0	5.0
	4.4	5.0
	44.6	5.0

CONTROL LOT	H 22410-13	
	Mean ²	± Range ³
	5.10	0.22
	95.4	5.0
	13.0	4.0
	48.7	3.5
	537	60
	12.5	1.5
	20.6	1.8
	16.4	0.5
	32.2	1.5
	33.7	3.0
	13.1	1.5
	0.8	1.0
	6.7	2.5
	63.8	5.0
	3.7	5.0
	32.5	5.0

- 1
- bg Трайност на отворен флакон 14 дни (bg)
 - hr Stabilitnost otvorene bočice 14 dani (hr)
 - cs Stabilita otevřené ampulky 14 dní (cs)
 - da Åben rør stabilitet 14 dage (da)
 - et Stabiilsus avatud vialis 14 päevade (et)
 - fr Stabilité en flacon ouvert 14 jour (fr)
 - de Stabilität geöffneter Flaschen 14 tage (de)
 - el Διατήρηση δείγματος μετά την αποσφράγιση 14 ημέρα (el)
 - it Stabilità della fiala aperta 14 giorni (it)
 - lv Atvērt flakonu stabilitāti 14 dienas (lv)
 - lt Stabilumas atidarius buteliuką 14 dienas (lt)
 - no Åpen rør/glass stabilitet 14 dager (no)
 - pl Trwałość po otwarciu fiołki 14 Liczba dni (pl)
 - pt Estabilidade após abertura do frasco 14 dias (pt)
 - ro Termenul de valabilitate al fiolei desfăcute 14 zile (ro)
 - ru Стабильность вскрытого флакона 14 дней (ru)
 - sr Proizvod stabilan nakon otvaranja pakovanja 14 dana (sr)
 - sk Stabilita otvorenej injekčnej liekovky 14 dní (sk)
 - sl Stabilita otvorenej injekčnej liekovky 14 dni (sl)
 - es Estabilidad de la cápsula abierta 14 días (es)
 - sv Hållbarhet för öppnad flaska 14 dagar (sv)
 - tr Açık şişe stabilitesi 14 gün (tr)

- 2
- Средно
 - Sredstvo
 - Průměr
 - Middelverdi
 - Keskmine
 - Moyenne
 - Mittelwert
 - Μέση
 - Media
 - Nozīmē
 - Vidurkis
 - Middelverdi
 - Średnia
 - Média
 - Medie
 - Среднее
 - Srednja vrednost
 - Priemer
 - Povprečno
 - Media
 - Medelvärde
 - Ortalama

- 3
- Обхват
 - Raspon
 - Rozsah
 - Område
 - Vahemik
 - Intervalle
 - Bereich
 - Αναμενόμενο εύρος α
 - Intervallo
 - Diapazons
 - Intervalas
 - Område
 - Zakres
 - Intervalo
 - Interval
 - диапазон
 - Orpsæg
 - Rozsah
 - Območje
 - Intervalo
 - Intervall
 - Aralığı

For Swelab Alfa / Alfa Plus with barcode reader

Alfa:

1. In the Main Menu go to "Q/C Menu" "Enter Con/Cal" "Control/Calibrator Assay Value Input"
2. Scan in each blood calibrator/control column by column, sequence 1 to 9

Alfa Plus:

1. In the Main Menu go to "Quality Control" "Input Assays"
2. Scan in each blood calibrator / control column by column, sequence 1 to 9

Instr.: Swelab Alfa
 Lot: 22410-11
 Prod.: Con-Diff Low
 ID: 2241011+
 Exp.: 3-Mar-2025

Instr.: Swelab Alfa
 Lot: 22410-12
 Prod.: Con-Diff Norm
 ID: 2241012+
 Exp.: 3-Mar-2025

Instr.: Swelab Alfa
 Lot: 22410-13
 Prod.: Con-Diff High
 ID: 2241013+
 Exp.: 6-Mar-2025

