

Certificate of Analysis

Product SWELAB LUMI-D DILUENT (20L)
Product No 1504514
Lot No 2024083001
Manufacturing Date 2024-08-30
Expiration Date 2026-08-29

Test	Specification
pH	7,1 – 7,5
Osmolality (mOsm/kg)	290 – 310
Conductivity (mS/cm)	17,3 – 18,3
Particles PLT/RBC channel (x 10 ⁹ /L)	≤ 5

Actual values meet specification.

Date: 2024-11-27



Elin Andersson
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

