a Boule Diagnostics, AB company



CERTIFICATE OF ANALYSIS

Product: Quintus 5-Part Lyse

Lot No: 2410-301

Product Part #: 501-291

Exp. Date: 2026-04-30

Manufacturing Date: 2024-10-17

Quality Testing Date: 2024-11-21

Parameter	Specification	Result	Pass/Fail
pH (units):	2.7 – 3.0	3.0	Pass
Osmolality (mOsm/kg):	35 – 44	37	Pass
Conductivity (µS/cm):	1,117 – 1,541	1,217	Pass

The product has been manufactured in compliance with Quality Systems Regulations and has met all current release specifications.

Quality Assurance

Date