

Certificate of Analysis

Product EXIGO LYSE, RFID (1.9 L)
Product No 1504500
Lot No 2412-267
Manufacturing Date 2024-12-09
Expiration Date 2027-12-08

Test	Specification
pH	6.95 – 7.25
Osmolality (mOsm/kg)	195 - 225
Conductivity (mS/cm)	12.10 – 12.80
CBC Performance *	Passes test
Particles (PLT/RBC channel)	≤ 5

* Performance test using calibrated Blood cell counter at Boule Medical reference laboratory. All parameters except mid cells are tested. Accepted reference 95 % to 105 % of expected control blood sample values.

Actual values meet specifications.

Date:2024-12-12



Ulrika Miefalk
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

