

# Boule Con-Diff

CONTROL

IVD



Control	Lot	
Tri-Level	22411-1K	2025-04-07
Low	22411-11	2025-04-07
Normal	22411-12	2025-04-07
High	22411-13	2025-04-10

Open vial stability 14 days <sup>1</sup>

MEDONIC		CONTROL	L
Medonic M-series/M32		LOT	22411-11
		Mean <sup>2</sup>	± Range <sup>3</sup>
RBC	10 <sup>12</sup> /L	2.26	0.12
MCV	fL	65.0	5.0
RDW	%	14.2	4.0
Hct	%	14.7	2.5
Plt	10 <sup>9</sup> /L	85	15
MPV	fL	10.9	1.5
WBC	10 <sup>9</sup> /L	3.4	0.3
Hgb	g/dL	5.5	0.3
MCH	pg	24.3	1.5
MCHC	g/dL	37.4	3.0
Lymph#	10 <sup>9</sup> /L	1.5	0.4
Mid#	10 <sup>9</sup> /L	0.2	0.3
Gran#	10 <sup>9</sup> /L	1.7	0.8
Lymph	%	43.4	7.0
Mid	%	7.2	5.0
Gran	%	49.4	7.0

CONTROL	N	
LOT	22411-12	
	Mean <sup>2</sup>	± Range <sup>3</sup>
	4.21	0.18
	84.2	5.0
	13.6	4.0
	35.4	3.0
	240	30
	11.2	1.5
	8.6	0.6
	12.5	0.4
	29.7	1.5
	35.3	3.0
	4.1	0.6
	0.6	0.5
	3.9	1.0
	47.4	5.0
	7.2	5.0
	45.4	5.0

CONTROL	H	
LOT	22411-13	
	Mean <sup>2</sup>	± Range <sup>3</sup>
	5.12	0.22
	94.0	5.0
	13.5	4.0
	48.1	3.5
	529	60
	11.7	1.5
	21.0	1.8
	16.5	0.5
	32.2	1.5
	34.3	3.0
	12.8	1.5
	1.2	1.0
	7.0	2.5
	60.9	5.0
	5.8	5.0
	33.3	5.0

1  
 bg Трайност на отворен флакон 14 дни (bg)  
 hr Stabilitnost otvorene bočice 14 dani (hr)  
 cs Stabilita otevřené ampulky 14 dní (cs)  
 da Åben rør stabilitet 14 dage (da)  
 et Stabilsus avatud vialis 14 päevade (et)  
 fr Stabilité en flacon ouvert 14 jour (fr)  
 de Stabilität geöffneter Flaschen 14 tage (de)  
 el Διατήρηση δείγματος μετά την αποσφράγιση 14 ημέρα (el)  
 it Stabilità della fiala aperta 14 giorni (it)  
 lv Atvērt flakonu stabilitāti 14 dienas (lv)  
 lt Stabulumas atidarius buteliuką 14 dienas (lt)  
 no Åpen rør/glass stabilitet 14 dager (no)  
 pl Trwałość po otwarciu fiołki 14 Liczba dni (pl)  
 pt Estabilidade após abertura do frasco 14 dias (pt)  
 ro Termenul de valabilitate al fiolei desfăcute 14 zile (ro)  
 ru Стабильность вскрытого флакона 14 дней (ru)  
 sr Proizvod stabilan nakon otvaranja pakovanja 14 dana (sr)  
 sk Stabilita otvorenej injekčnej liekovky 14 dní (sk)  
 sl Stabilita otvorenej injekčnej liekovky 14 dni (sl)  
 es Estabilidad de la cápsula abierta 14 días (es)  
 sv Hållbarhet för öppnad flaska 14 dagar (sv)  
 tr Açık şişe stabilitesi 14 gün (tr)

2  
 Средно  
 Sredstvo  
 Průměr  
 Middelværdi  
 Keskmine  
 Moyenne  
 Mittelwert  
 Μέση  
 Media  
 Nozīmē  
 Vidurkis  
 Middelværdi  
 Średnia  
 Média  
 Media  
 Среднее  
 Srednja vrednost  
 Priemer  
 Povprečno  
 Media  
 Medelvärde  
 Ortalama

3  
 Обхват  
 Raspon  
 Rozsah  
 Område  
 Vahemik  
 Intervalle  
 Bereich  
 Αναμενόμενο εύρος α  
 Intervallo  
 Diapazons  
 Intervalas  
 Område  
 Zakres  
 Intervalo  
 Interval  
 диапазон  
 Orseg  
 Rozsah  
 Območje  
 Intervall  
 Aralık

# For Medonic M-series / M32 with barcode reader

## M-Series:

1. In the Main Menu go to "Q/C Menu" "Enter Con/Cal" "Control/Calibrator Assay Value Input"
2. Scan in each blood calibrator/control column by column, sequence 1 to 9

## M32:

1. In the Main Menu go to "Quality Control" "Input Assays"
2. Scan in each blood calibrator / control column by column, sequence 1 to 9

Instr.: M-series  
 Lot: 22411-11  
 Prod.: Con-Diff Low  
 ID: 2241111+  
 Exp.: 7-Apr-2025

Instr.: M-series  
 Lot: 22411-12  
 Prod.: Con-Diff Norm  
 ID: 2241112+  
 Exp.: 7-Apr-2025

Instr.: M-series  
 Lot: 22411-13  
 Prod.: Con-Diff High  
 ID: 2241113+  
 Exp.: 10-Apr-2025



# Boule Con-Diff

CONTROL

IVD



Control	Lot	
Tri-Level	22411-1K	2025-04-07
Low	22411-11	2025-04-07
Normal	22411-12	2025-04-07
High	22411-13	2025-04-10

Open vial stability 14 days <sup>1</sup>

## SWELAB

CONTROL		L	
LOT		22411-11	
		Mean <sup>2</sup>	± Range <sup>3</sup>
Swelab Alfa/Alfa Plus	10 <sup>12</sup> /L	2.28	0.12
RBC	fL	66.6	5.0
MCV	fL	13.7	4.0
RDW	%	15.2	2.5
Hct	%	85	15
Plt	10 <sup>9</sup> /L	11.9	1.5
MPV	fL	3.4	0.3
WBC	10 <sup>9</sup> /L	5.5	0.3
Hgb	g/dL	24.1	1.5
MCH	pg	36.2	3.0
MCHC	g/dL	1.6	0.4
Lymph#	10 <sup>9</sup> /L	0.2	0.3
Mid#	10 <sup>9</sup> /L	1.6	0.8
Gran#	10 <sup>9</sup> /L	48.0	7.0
Lymph	%	4.8	5.0
Mid	%	47.2	7.0
Gran	%		

CONTROL		N	
LOT		22411-12	
		Mean <sup>2</sup>	± Range <sup>3</sup>
		4.22	0.18
		85.7	5.0
		12.6	4.0
		36.2	3.0
		243	30
		12.2	1.5
		8.5	0.6
		12.5	0.4
		29.6	1.5
		34.5	3.0
		4.3	0.6
		0.4	0.5
		3.8	1.0
		51.3	5.0
		4.5	5.0
		44.2	5.0

CONTROL		H	
LOT		22411-13	
		Mean <sup>2</sup>	± Range <sup>3</sup>
		5.14	0.22
		96.0	5.0
		12.6	4.0
		49.3	3.5
		547	60
		12.8	1.5
		20.8	1.8
		16.5	0.5
		32.1	1.5
		33.5	3.0
		13.2	1.5
		0.8	1.0
		6.8	2.5
		63.7	5.0
		3.7	5.0
		32.6	5.0

- 1
- bg Трайност на отворен флакон 14 дни (bg)
  - hr Stabilnost otvorene bočice 14 dani (hr)
  - cs Stabilita otevřené ampulky 14 dní (cs)
  - da Åben rør stabilitet 14 dage (da)
  - et Stabiilsus avatud vialis 14 päevade (et)
  - fr Stabilité en flacon ouvert 14 jour (fr)
  - de Stabilität geöffneter Flaschen 14 tage (de)
  - el Διατήρηση δείγματος μετά την αποσφράγιση 14 ημέρα (el)
  - it Stabilità della fiala aperta 14 giorni (it)
  - lv Atvērt flakonu stabilitāti 14 dienas (lv)
  - lt Stabilumas atidarius buteliuką 14 dienas (lt)
  - no Åpen rør/glass stabilitet 14 dager (no)
  - pl Trwałość po otwarciu fiołki 14 Liczba dni (pl)
  - pt Estabilidade após abertura do frasco 14 dias (pt)
  - ro Termenul de valabilitate al fiolei desfăcute 14 zile (ro)
  - ru Стабильность вскрытого флакона 14 дней (ru)
  - sr Proizvod stabilan nakon otvaranja pakovanja 14 dana (sr)
  - sk Stabilita otvorenej injekčnej liekovky 14 dní (sk)
  - sl Stabilita otvorenej injekčnej liekovky 14 dni (sl)
  - es Estabilidad de la cápsula abierta 14 días (es)
  - sv Hållbarhet för öppnad flaska 14 dagar (sv)
  - tr Açık şişe stabilitesi 14 gün (tr)

- 2
- Средно
  - Sredstvo
  - Průměr
  - Middelverdi
  - Keskmine
  - Moyenne
  - Mittelwert
  - Μέση
  - Media
  - Nozimë
  - Vidurkis
  - Middelverdi
  - Średnia
  - Média
  - Medie
  - Среднее
  - Srednja vrednost
  - Priemer
  - Povprečno
  - Media
  - Medelvärde
  - Ortalama

- 3
- Обхват
  - Raspon
  - Rozsah
  - Område
  - Vahemik
  - Intervalle
  - Bereich
  - Αναμενόμενο εύρος α
  - Intervallo
  - Diapazons
  - Intervalas
  - Område
  - Zakres
  - Intervalo
  - Interval
  - диапазон
  - Orseg
  - Rozsah
  - Območje
  - Intervalo
  - Intervall
  - Aralığı

## For Swelab Alfa / Alfa Plus with barcode reader

### Alfa:

1. In the Main Menu go to "Q/C Menu" "Enter Con/Cal" "Control/Calibrator Assay Value Input"
2. Scan in each blood calibrator/control column by column, sequence 1 to 9

### Alfa Plus:

1. In the Main Menu go to "Quality Control" "Input Assays"
2. Scan in each blood calibrator / control column by column, sequence 1 to 9

Instr.: Swelab Alfa  
Lot: 22411-11  
Prod.: Con-Diff Low  
ID: 2241111+  
Exp.: 7-Apr-2025

Instr.: Swelab Alfa  
Lot: 22411-12  
Prod.: Con-Diff Norm  
ID: 2241112+  
Exp.: 7-Apr-2025

Instr.: Swelab Alfa  
Lot: 22411-13  
Prod.: Con-Diff High  
ID: 2241113+  
Exp.: 10-Apr-2025

