


# Certificate of Analysis

**Product** Quintus 5-part Lyse 5L  
**Product No** 1504292  
**Lot No** 2412-269  
**Manufacturing Date** 2024-12-10  
**Expiration Date** 2026-06-09

Test	Specification
pH	2,7 – 3,0
Osmolality (mOsm/kg)	35 – 44
Conductivity ( $\mu$ S/cm)	1117 – 1541
Background WBC ( $\times 10^{12}/L$ )	$\leq 0.1$

Actual values meet specification.

Date: 2025-01-31



*Caroline Sparf*  
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.  
All information in this document is property of Boule Medical AB.

