

CERTIFICATE OF ANALYSIS

Product: Medonic™ M-Series Lyse

Lot No: 2412-003

Product Part #: 501-212


Exp. Date: 2027-11-30

Manufacturing Date: 2024-12-18

Quality Testing Date: 2025-01-02

Parameter	Specification	Result	Pass/Fail
pH (units):	6.80 – 7.20	7.13	Pass
Osmolality (mOsm/kg):	220 – 240	233	Pass
Conductivity (mS/cm):	14.00 – 15.00	14.18	Pass
Functional Test	Pass	Pass	Pass

The product has been manufactured in compliance with Quality Systems Regulations and has met all current release specifications.



Quality Assurance



Date

