

# CERTIFICATE OF ANALYSIS

**Product:** Boule Cleaning Kit

**Kit Lot No:** 2147

**Product Part #:** 501-036

**Kit Exp. Date:** 2026-12-31

<b>Part Number:</b> 503-085			
<b>Boule Detergent Cleaner Lot # 2144</b>			
<b>Manufacturing Date:</b> 2024-12-09		<b>Quality Testing Date:</b> 2024-12-13	
Parameter	Specification	Result	Pass/Fail
pH (units):	7.60 – 8.40	8.25	Pass

<b>Part Number:</b> 503-086			
<b>Boule Hypochlorite (2%) Cleaner Lot # 2145</b>			
<b>Manufacturing Date:</b> 2024-12-12		<b>Quality Testing Date:</b> 2024-12-13	
Parameter	Specification	Result	Pass/Fail
pH (units):	11.80 – 13.00	12.94	Pass

<b>Part Number:</b> 503-087			
<b>Boule Enzymatic Cleaner Lot # 2146</b>			
<b>Manufacturing Date:</b> 2024-12-06		<b>Quality Testing Date:</b> 2024-12-13	
Parameter	Specification	Result	Pass/Fail
pH (units):	7.60 – 8.40	8.03	Pass

The product has been manufactured in compliance with Quality Systems Regulations and has met all current release specifications.

  
\_\_\_\_\_  
Quality and Regulatory

01/03/25  
\_\_\_\_\_  
Date

