a Boule Diagnostics, AB company



CERTIFICATE OF ANALYSIS

Product: Quintus 5-Part Stopper Lot No: 2501-203

Product Part #: 501-350 **Exp. Date:** 2027-12-31

Manufacturing Date: 2025-01-24

Quality Testing Date: 2025-02-05

Parameter	Specification	Result	Pass/Fail
pH (units):	10.10 - 11.20	10.89	Pass
Osmolality (mOsm/kg):	830 – 860	843	Pass
Conductivity (mS/cm):	42.7 – 43.3	42.8	Pass
Functional Test	Pass	Pass	Pass

The product has been manufactured in compliance with Quality Systems Regulations and has met all current release specifications.

Quality Assurance

Date