

CERTIFICATE OF ANALYSIS

Product: Medonic™ M-Series Lyse

Lot No: 2501-001

Product Part #: 501-212

Exp. Date: 2027-12-31

Manufacturing Date: 2025-01-27

Quality Testing Date: 2025-02-06

| Parameter | Specification | Result | Pass/Fail |
|-----------------------|---------------|--------|-----------|
| pH (units): | 6.80 – 7.20 | 7.09 | Pass |
| Osmolality (mOsm/kg): | 220 – 240 | 230 | Pass |
| Conductivity (mS/cm): | 14.00 – 15.00 | 14.12 | Pass |
| Functional Test | Pass | Pass | Pass |

The product has been manufactured in compliance with Quality Systems Regulations and has met all current release specifications.



Quality Assurance



Date

