

CERTIFICATE OF ANALYSIS

Product: Quintus 5-Part Stopper

Lot No: 2502-202

Product Part #: 501-350

Exp. Date: 2028-01-31

Manufacturing Date: 2025-02-19


Quality Testing Date: 2025-02-26

Parameter	Specification	Result	Pass/Fail
pH (units):	10.10 – 11.20	11.02	Pass
Osmolality (mOsm/kg):	830 – 860	846	Pass
Conductivity (mS/cm):	42.7 – 43.3	42.8	Pass
Functional Test	Pass	Pass	Pass

The product has been manufactured in compliance with Quality Systems Regulations and has met all current release specifications.



Quality Assurance



Date

