

Certificate of Analysis

Product	MEDONIC M51-D DILUENT (20L)
Product No	1504510
Lot No	2025011801
Manufacturing Date	2025-01-18
Expiration Date	2027-01-17

Test	Specification
pH	7,1 – 7,5
Osmolality (mOsm/kg)	290 – 310
Conductivity (mS/cm)	17,3 – 18,3
Particles PLT/RBC channel (x 10 ⁹ /L)	≤ 5

Actual values meet specification.

Date: 2025-04-02



Ulrika Miefalk
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

