

Certificate of Analysis

Product MEDONIC M51-L2 LYSE (500 mL)
Product No 1504512
Lot No 2024122701
Manufacturing Date 2024-12-27
Expiration Date 2026-12-26

Test	Specification
pH	4,0 – 4,4
CBC Performance test*	Passes test
Background WBC (x 10 ⁹ /L)	≤ 0,2

*Performance testing at a reference laboratory. WBC, HGB and WBC scattergrams tested and approved.

Actual values meet specification.

Date: 2025-04-02



Elin Andersson
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

