a Boule Diagnostics, AB company



1998

CERTIFICATE OF ANALYSIS

Product: Quintus 5-Part Lyse Lot No: 2406-301

Product Part #: 501-291 Exp. Date: 2026-01-31

Manufacturing Date: 2024-07-22

Quality Testing Date: 2024-08-23

Parameter	Specification	Result	Pass/Fail
pH (units):	2.7 – 3.0	2.9	Pass
Osmolality (mOsm/kg):	35 – 44	36	Pass
Conductivity (µS/cm):	1,117 – 1,541	1,231	Pass

The product has been manufactured in compliance with Quality Systems Regulations and has met all current release specifications.

Regulatory and Compliance

Date