

CERTIFICATE OF ANALYSIS

Product: Quintus 5-Part Diluent

Lot No: 2506-201

Product Part #: 501-292

Exp. Date: 2027-05-31

Manufacturing Date: 2025-06-17

Quality Testing Date: 2025-06-20

Parameter	Specification	Result	Pass/Fail
pH (units):	6.7 – 7.1	7.1	Pass
Osmolality (mOsm/kg):	320 – 350	332	Pass
Conductivity (mS/cm):	18.50 – 19.50	18.54	Pass
Background ($\times 10^3/\mu\text{L}$)	0 – 8	3	Pass
Functional Test	Pass	Pass	Pass

The product has been manufactured in compliance with Quality Systems Regulations and has met all current release specifications.



Quality and Regulatory

06/30/25

Date

