

# Certificate of Analysis

<b>Product</b>	Exigo System Cleaner (1.9 L)
<b>Product No</b>	1504203
<b>Lot No</b>	2203-678
<b>Manufacturing Date</b>	2022-03-07
<b>Expiration Date</b>	2025-03-07


<b>Test</b>	<b>Specification</b>
pH	7,00 – 8,00
Conductivity (mS)	15,25 – 17,00
Particles (PLT/RBC channel)	≤ 30

Actual values meet specifications.

Date: 2022-03-17



*Lubna Elabbas*  
Quality Assurance



This product is intended for Laboratory use only unless otherwise specifically stated.  
All information in this document is property of Boule Medical AB.

