

Certificate of Analysis

Product Boule Micropipette Plastic EDTA 10x100 pcs
Boule Micropipette Plastic EDTA 1x100 pcs
Product No 1070030
1070039
Lot No 22274525
Manufacturing Date 2022-07-26
Expiration Date 2025-06-28

Parameter	Specification
Volume 20 µL	CV ≤ 1,0%
WBC background	≤ 0,2 x 10 ⁹ /L
RBC background	≤ 0,03 x 10 ¹² /L
HGB background	≤ 0,2 x g/dL
PLT background	≤ 10 x 10 ⁹ /L

Actual value meets specification.

Date: 2022-12-13



Ulrika Miefalk
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

