

Certificate of Analysis

Product	Boule Micropipette Plastic EDTA 10x100 pcs Boule Micropipette Plastic EDTA 1x100 pcs
Product No	1070030 1070039
Lot No	224045M9
Manufacturing Date	2022-12-07
Expiration Date	2025-10-28

Parameter	Specification
Volume 20 μ L	$CV \leq 1,0\%$
WBC background	$\leq 0,2 \times 10^9/L$
RBC background	$\leq 0,03 \times 10^{12}/L$
HGB background	$\leq 0,2 \times g/dL$
PLT background	$\leq 10 \times 10^9/L$

Actual value meets specification.

Date: 2023-01-17



Sofia Karlström
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

