

# Boule Cal

CAL

IVD



Calibrator

Lot	
22503-34	2025-06-15

Open vial stability 5 days <sup>1</sup>

## Medonic

Parameter <sup>2</sup>	RBC 10 <sup>12</sup> /L	MCV fL	Plt 10 <sup>9</sup> /L	MPV fL	WBC 10 <sup>9</sup> /L	Hgb g/dL
Medonic M-series / M32	4.19	80.8	242	11.3	8.3	12.2
Range ± <sup>3</sup>	0.10	2.0	15	1.5	0.3	0.2

## Swelab

Parameter <sup>2</sup>	RBC 10 <sup>12</sup> /L	MCV fL	Plt 10 <sup>9</sup> /L	MPV fL	WBC 10 <sup>9</sup> /L	Hgb g/dL
Swelab Alfa / Alfa Plus	4.19	81.4	247	11.9	8.3	12.2
Range ± <sup>3</sup>	0.10	2.0	15	1.5	0.3	0.2

	1	2	3
bg	Трайност на отворен флакон	5 дни	Обхват
hr	Stabilnost otvorene bočice	5 dani	Raspon
cs	Stabilita otevřené ampulky	5 dní	Rozsah
da	Åben r�r stabilitet	5 dage	Omr�de
et	Stabiilsus avatud viaalis	5 p�evade arv	Vahemik
fr	Stabilit� en flacon ouvert	5 jour	Intervalle
de	Stabilit�t ge�ffneter Flaschen	5 tage	Bereich
el	Διατ�ρηση δ�ιγματος μετ� την αποσφ�ραγιση	5 ημ�ρα	Αναμενόμενο ε�ρος α
it	Stabilit� della fiala aperta	5 giorni	Intervallo
lv	Atv�rt flakonu stabilit�ti	5 dienas	Diapazons
lt	Stabilumas atidarius buteliuka	5 d.	Intervalas
no	�pen r�r/glass stabilitet	5 dager	Omr�de
pl	Trwa�o�c po otworzeniu fiolki	5 Liczba dni	Zakres
pt	Estabilidade ap�s abertura do frasco	5 dias	Intervalo
ro	Termenul de valabilitate al fiolei desfacute	5 zile	Interval
ru	Стабильность вскрытого флакона	5 дней	диапазон
sr	Proizvod stabilan nakon otvaranja pakovanja	5 dana	Opseg
sk	Stabilita otvorenej injek�nej liekovky	5 dn�	Rozsah
sl	Stabilnost odprte stekleni�ke	5 dni	Parameter
es	Estabilidad de la c�psula abierta	5 dias	Par�metro
sv	H�llbarhet f�r �ppnad flaska	5 dagar	Parameter
tr	A�ık ŐiŐe stabilitesi	5 g�n	Parametre

[www.boule.com](http://www.boule.com)Boule Medical AB  
Domnarvsgatan 4  
SE-163 53 Sp nga, Sweden

25123-6

1504055

# For Medonic M-series / M32 with barcode reader

## M-Series:

1. In the Main Menu go to "Q/C Menu" "Enter Con/Cal" "Control/Calibrator Assay Value Input"
2. Scan in each blood calibrator/control column by column, sequence 1 to 9

## M32:

1. In the Main Menu go to "Quality Control" "Input Assays"
2. Scan in each blood calibrator / control column by column, sequence 1 to 9

Instr.: M-series

Lot: 22503-34

Prod.: Cal

ID: 2250334+

Exp.: 15-Jun-2025



## For Swelab Alfa / Alfa Plus with barcode reader

### Alfa:

1. In the Main Menu go to "Q/C Menu" "Enter Con/Cal" "Control/Calibrator Assay Value Input"
2. Scan in each blood calibrator/control column by column, sequence 1 to 9

### Alfa Plus:

1. In the Main Menu go to "Quality Control" "Input Assays"
2. Scan in each blood calibrator / control column by column, sequence 1 to 9

Instr.: Swelab Alfa

Lot: 22503-34

Prod.: Cal

ID: 2250334+

Exp.: 15-Jun-2025

