

Boule Con-Diff

CONTROL

IVD



| Control | Lot | |
|-----------|----------|------------|
| Tri-Level | 22412-1K | 2025-05-05 |
| Low | 22412-11 | 2025-05-05 |
| Normal | 22412-12 | 2025-05-05 |
| High | 22412-13 | 2025-05-07 |

Open vial stability 14 days ¹

| MEDONIC | | CONTROL | L |
|----------------------|---------------------|-------------------|----------------------|
| Medonic M-series/M32 | | LOT | 22412-11 |
| | | Mean ² | ± Range ³ |
| RBC | 10 ¹² /L | 2.26 | 0.12 |
| MCV | fL | 64.8 | 5.0 |
| RDW | % | 14.6 | 4.0 |
| Hct | % | 14.6 | 2.5 |
| Plt | 10 ⁹ /L | 82 | 15 |
| MPV | fL | 11.3 | 1.5 |
| WBC | 10 ⁹ /L | 3.4 | 0.3 |
| Hgb | g/dL | 5.5 | 0.3 |
| MCH | pg | 24.3 | 1.5 |
| MCHC | g/dL | 37.7 | 3.0 |
| Lymph# | 10 ⁹ /L | 1.5 | 0.4 |
| Mid# | 10 ⁹ /L | 0.2 | 0.3 |
| Gran# | 10 ⁹ /L | 1.7 | 0.8 |
| Lymph | % | 42.8 | 7.0 |
| Mid | % | 6.3 | 5.0 |
| Gran | % | 50.9 | 7.0 |

| CONTROL | N | |
|---------|-------------------|----------------------|
| LOT | 22412-12 | |
| | Mean ² | ± Range ³ |
| | 4.22 | 0.18 |
| | 83.3 | 5.0 |
| | 13.0 | 4.0 |
| | 35.2 | 3.0 |
| | 242 | 30 |
| | 11.6 | 1.5 |
| | 8.8 | 0.6 |
| | 12.4 | 0.4 |
| | 29.4 | 1.5 |
| | 35.2 | 3.0 |
| | 4.1 | 0.6 |
| | 0.7 | 0.5 |
| | 4.0 | 1.0 |
| | 46.6 | 5.0 |
| | 7.6 | 5.0 |
| | 45.8 | 5.0 |

| CONTROL | H | |
|---------|-------------------|----------------------|
| LOT | 22412-13 | |
| | Mean ² | ± Range ³ |
| | 5.01 | 0.22 |
| | 92.1 | 5.0 |
| | 13.0 | 4.0 |
| | 46.1 | 3.5 |
| | 514 | 60 |
| | 11.8 | 1.5 |
| | 20.8 | 1.8 |
| | 16.1 | 0.5 |
| | 32.1 | 1.5 |
| | 34.9 | 3.0 |
| | 12.6 | 1.5 |
| | 1.3 | 1.0 |
| | 6.9 | 2.5 |
| | 60.5 | 5.0 |
| | 6.2 | 5.0 |
| | 33.3 | 5.0 |

- 1
- bg Трайност на отворен флакон 14 дни (bg)
 - hr Stabilnost otvorene bočice 14 dani (hr)
 - cs Stabilita otevřené ampulky 14 dní (cs)
 - da Åben rør stabilitet 14 dage (da)
 - et Stabiilsus avatud vialis 14 päevade (et)
 - fr Stabilité en flacon ouvert 14 jour (fr)
 - de Stabilität geöffneter Flaschen 14 tage (de)
 - el Διατήρηση δείγματος μετά την αποσφράγιση 14 ημέρα (el)
 - it Stabilità della fiala aperta 14 giorni (it)
 - lv Atvērt flakonu stabilitāti 14 dienas (lv)
 - lt Stabilumas atidarius buteliuką 14 dienas (lt)
 - no Åpen rør/glass stabilitet 14 dager (no)
 - pl Trwałość po otwarciu fiołki 14 Liczba dni (pl)
 - pt Estabilidade após abertura do frasco 14 dias (pt)
 - ro Termenul de valabilitate al fiolei desfăcute 14 zile (ro)
 - ru Стабильность вскрытого флакона 14 дней (ru)
 - sr Proizvod stabilan nakon otvaranja pakovanja 14 dana (sr)
 - sk Stabilita otvorenej injekčnej liekovky 14 dní (sk)
 - sl Stabilita otvorenej injekčnej liekovky 14 dni (sl)
 - es Estabilidad de la cápsula abierta 14 días (es)
 - sv Hållbarhet för öppnad flaska 14 dagar (sv)
 - tr Açık şişe stabilitesi 14 gün (tr)

- 2
- Средно
 - Sredstvo
 - Průměr
 - Middelværdi
 - Keskmine
 - Moyenne
 - Mittelwert
 - Μέση
 - Media
 - Nozīmē
 - Vidurkis
 - Middelverdi
 - Średnia
 - Média
 - Medie
 - Среднее
 - Srednja vrednost
 - Priemer
 - Povprečno
 - Media
 - Medelvärde
 - Ortalama

- 3
- Обхват
 - Raspon
 - Rozsah
 - Område
 - Vahemik
 - Intervalle
 - Bereich
 - Αναμενόμενο εύρος α
 - Intervallo
 - Diapazons
 - Intervalas
 - Område
 - Zakres
 - Intervalo
 - Interval
 - диапазон
 - Opseg
 - Rozsah
 - Območje
 - Interval
 - Intervall
 - Aralığı

For Medonic M-series / M32 with barcode reader

M-Series:

1. In the Main Menu go to "Q/C Menu" "Enter Con/Cal" "Control/Calibrator Assay Value Input"
2. Scan in each blood calibrator/control column by column, sequence 1 to 9

M32:

1. In the Main Menu go to "Quality Control" "Input Assays"
2. Scan in each blood calibrator / control column by column, sequence 1 to 9

Instr.: M-series
 Lot: 22412-11
 Prod.: Con-Diff Low
 ID: 2241211+
 Exp.: 5-May-2025

Instr.: M-series
 Lot: 22412-12
 Prod.: Con-Diff Norm
 ID: 2241212+
 Exp.: 5-May-2025

Instr.: M-series
 Lot: 22412-13
 Prod.: Con-Diff High
 ID: 2241213+
 Exp.: 7-May-2025



Boule Con-Diff

CONTROL

IVD



| Control | Lot | |
|-----------|----------|------------|
| Tri-Level | 22412-1K | 2025-05-05 |
| Low | 22412-11 | 2025-05-05 |
| Normal | 22412-12 | 2025-05-05 |
| High | 22412-13 | 2025-05-07 |

Open vial stability 14 days ¹

| SWELAB | | CONTROL | L |
|-----------------------|---------------------|-------------------|----------------------|
| Swelab Alfa/Alfa Plus | | LOT | 22412-11 |
| | | Mean ² | ± Range ³ |
| RBC | 10 ¹² /L | 2.28 | 0.12 |
| MCV | fL | 66.1 | 5.0 |
| RDW | % | 14.0 | 4.0 |
| Hct | % | 15.1 | 2.5 |
| Plt | 10 ⁹ /L | 82 | 15 |
| MPV | fL | 12.1 | 1.5 |
| WBC | 10 ⁹ /L | 3.4 | 0.3 |
| Hgb | g/dL | 5.6 | 0.3 |
| MCH | pg | 24.6 | 1.5 |
| MCHC | g/dL | 37.1 | 3.0 |
| Lymph# | 10 ⁹ /L | 1.6 | 0.4 |
| Mid# | 10 ⁹ /L | 0.1 | 0.3 |
| Gran# | 10 ⁹ /L | 1.7 | 0.8 |
| Lymph | % | 46.9 | 7.0 |
| Mid | % | 3.9 | 5.0 |
| Gran | % | 49.2 | 7.0 |

| CONTROL | N | |
|---------|-------------------|----------------------|
| LOT | 22412-12 | |
| | Mean ² | ± Range ³ |
| | 4.23 | 0.18 |
| | 84.2 | 5.0 |
| | 12.1 | 4.0 |
| | 35.6 | 3.0 |
| | 247 | 30 |
| | 12.5 | 1.5 |
| | 8.7 | 0.6 |
| | 12.5 | 0.4 |
| | 29.6 | 1.5 |
| | 35.1 | 3.0 |
| | 4.5 | 0.6 |
| | 0.4 | 0.5 |
| | 3.8 | 1.0 |
| | 51.3 | 5.0 |
| | 4.5 | 5.0 |
| | 44.2 | 5.0 |

| CONTROL | H | |
|---------|-------------------|----------------------|
| LOT | 22412-13 | |
| | Mean ² | ± Range ³ |
| | 4.99 | 0.22 |
| | 93.4 | 5.0 |
| | 12.1 | 4.0 |
| | 46.6 | 3.5 |
| | 527 | 60 |
| | 12.7 | 1.5 |
| | 20.6 | 1.8 |
| | 16.0 | 0.5 |
| | 32.1 | 1.5 |
| | 34.3 | 3.0 |
| | 13.1 | 1.5 |
| | 0.7 | 1.0 |
| | 6.8 | 2.5 |
| | 63.6 | 5.0 |
| | 3.6 | 5.0 |
| | 32.8 | 5.0 |

- 1
- bg Трайност на отворен флакон 14 дни (bg)
 - hr Stabilitnost otvorene bočice 14 dani (hr)
 - cs Stabilita otevřené ampulky 14 dní (cs)
 - da Åben rør stabilitet 14 dage (da)
 - et Stabiilsus avatud vialis 14 päevade (et)
 - fr Stabilité en flacon ouvert 14 jour (fr)
 - de Stabilität geöffneter Flaschen 14 tage (de)
 - el Διατήρηση δείγματος μετά την αποσφράγιση 14 ημέρα (el)
 - it Stabilità della fiala aperta 14 giorni (it)
 - lv Atvērt flakonu stabilitāti 14 dienas (lv)
 - lt Stabilumas atidarius buteliuką 14 dienas (lt)
 - no Åpen rør/glass stabilitet 14 dager (no)
 - pl Trwałość po otwarciu fiołki 14 Liczba dni (pl)
 - pt Estabilidade após abertura do frasco 14 dias (pt)
 - ro Termenul de valabilitate al fiolei desfăcute 14 zile (ro)
 - ru Стабильность вскрытого флакона 14 дней (ru)
 - sr Proizvod stabilan nakon otvaranja pakovanja 14 dana (sr)
 - sk Stabilita otvorenej injekčnej liekovky 14 dní (sk)
 - sl Stabilita otvorenej injekčnej liekovky 14 dni (sl)
 - es Estabilidad de la cápsula abierta 14 días (es)
 - sv Hållbarhet för öppnad flaska 14 dagar (sv)
 - tr Açık şişe stabilitesi 14 gün (tr)

- 2
- Средно
 - Sredstvo
 - Průměr
 - Middelverdi
 - Keskmine
 - Moyenne
 - Mittelwert
 - Μέση
 - Media
 - Nozīmē
 - Vidurkis
 - Middelverdi
 - Średnia
 - Média
 - Medie
 - Среднее
 - Srednja vrednost
 - Priemer
 - Povprečno
 - Media
 - Medelvärde
 - Ortalama

- 3
- Обхват
 - Raspon
 - Rozsah
 - Område
 - Vahemik
 - Intervalle
 - Bereich
 - Αναμενόμενο εύρος α
 - Intervallo
 - Diapazons
 - Intervalas
 - Område
 - Zakres
 - Intervalo
 - Interval
 - диапазон
 - Orseg
 - Rozsah
 - Območje
 - Intervalo
 - Intervall
 - Aralığı

For Swelab Alfa / Alfa Plus with barcode reader

Alfa:

1. In the Main Menu go to "Q/C Menu" "Enter Con/Cal" "Control/Calibrator Assay Value Input"
2. Scan in each blood calibrator/control column by column, sequence 1 to 9

Alfa Plus:

1. In the Main Menu go to "Quality Control" "Input Assays"
2. Scan in each blood calibrator / control column by column, sequence 1 to 9

Instr.: Swelab Alfa
Lot: 22412-11
Prod.: Con-Diff Low
ID: 2241211+
Exp.: 5-May-2025

Instr.: Swelab Alfa
Lot: 22412-12
Prod.: Con-Diff Norm
ID: 2241212+
Exp.: 5-May-2025

Instr.: Swelab Alfa
Lot: 22412-13
Prod.: Con-Diff High
ID: 2241213+
Exp.: 7-May-2025

