

Boule Con-Diff

CONTROL

IVD



Open vial stability 14 days ¹

Control	Lot	
Tri-Level	--	--
Low	22412-41	2025-05-05
Normal	22412-42	2025-05-05
High	22412-43	2025-05-07

MEDONIC

Medonic M-series/M32		CONTROL	
		LOT	22412-41
		Mean ²	± Range ³
RBC	10 ¹² /L	2.26	0.12
MCV	fL	64.8	5.0
RDW	%	14.6	4.0
Hct	%	14.6	2.5
Plt	10 ⁹ /L	82	15
MPV	fL	11.3	1.5
WBC	10 ⁹ /L	3.4	0.3
Hgb	g/dL	5.5	0.3
MCH	pg	24.3	1.5
MCHC	g/dL	37.7	3.0
Lymph#	10 ⁹ /L	1.5	0.4
Mid#	10 ⁹ /L	0.2	0.3
Gran#	10 ⁹ /L	1.7	0.8
Lymph	%	42.8	7.0
Mid	%	6.3	5.0
Gran	%	50.9	7.0

CONTROL	LOT	N	
		22412-42	
		Mean ²	± Range ³
		4.22	0.18
		83.3	5.0
		13.0	4.0
		35.2	3.0
		242	30
		11.6	1.5
		8.8	0.6
		12.4	0.4
		29.4	1.5
		35.2	3.0
		4.1	0.6
		0.7	0.5
		4.0	1.0
		46.6	5.0
		7.6	5.0
		45.8	5.0

CONTROL	LOT	H	
		22412-43	
		Mean ²	± Range ³
		5.01	0.22
		92.1	5.0
		13.0	4.0
		46.1	3.5
		514	60
		11.8	1.5
		20.8	1.8
		16.1	0.5
		32.1	1.5
		34.9	3.0
		12.6	1.5
		1.3	1.0
		6.9	2.5
		60.5	5.0
		6.2	5.0
		33.3	5.0

1

bg	Трайност на отворен флакон 14 дни (bg)
hr	Stabilnost otvorene bočice 14 dani (hr)
cs	Stabilita otevřené ampulky 14 dní (cs)
da	Åben rør stabilitet 14 dage (da)
et	Stabiilsus avatud vialis 14 päevade (et)
fr	Stabilité en flacon ouvert 14 jour (fr)
de	Stabilität geöffneter Flaschen 14 tage (de)
el	Διατήρηση δείγματος μετά την αποσφράγιση 14 ημέρα (el)
it	Stabilità della fiala aperta 14 giorni (it)
lv	Atvērt flakonu stabilitāti 14 dienas (lv)
lt	Stabilumas atidarius buteliuką 14 dienas (lt)
no	Åpen rør/glass stabilitet 14 dager (no)
pl	Trwałość po otwarciu fiołki 14 Liczba dni (pl)
pt	Estabilidade após abertura do frasco 14 dias (pt)
ro	Termenul de valabilitate al fiolei desfăcute 14 zile (ro)
ru	Стабильность вскрытого флакона 14 дней (ru)
sr	Proizvod stabilan nakon otvaranja pakovanja 14 dana (sr)
sk	Stabilita otvorenej injekčnej liekovky 14 dní (sk)
sl	Stabilita otvorenej injekčnej liekovky 14 dni (sl)
es	Estabilidad de la cápsula abierta 14 días (es)
sv	Hållbarhet för öppnad flaska 14 dagar (sv)
tr	Açık şişe stabilitesi 14 gün (tr)

2

	Средно
	Sredstvo
	Průměr
	Middelværdi
	Keskmine
	Moyenne
	Mittelwert
	Μέση
	Media
	Nozīmē
	Vidurkis
	Middelverdi
	Średnia
	Média
	Medie
	Среднее
	Srednja vrednost
	Priemer
	Povprečno
	Media
	Medelvärde
	Ortalama

3

	Обхват
	Raspon
	Rozsah
	Område
	Vahemik
	Intervalle
	Bereich
	Αναμενόμενο εύρος α
	Intervallo
	Diapazons
	Intervalas
	Område
	Zakres
	Intervalo
	Interval
	диапазон
	Opseg
	Rozsah
	Območje
	Intervallo
	Intervall
	Aralığı

For Medonic M-series / M32 with barcode reader























M-Series:

1. In the Main Menu go to "Q/C Menu" "Enter Con/Cal" "Control/Calibrator Assay Value Input"
2. Scan in each blood calibrator/control column by column, sequence 1 to 9

M32:

1. In the Main Menu go to "Quality Control" "Input Assays"
2. Scan in each blood calibrator / control column by column, sequence 1 to 9

Instr.: M-series	Instr.: M-series	Instr.: M-series
Lot: 22412-41	Lot: 22412-42	Lot: 22412-43
Prod.: Con-Diff Low	Prod.: Con-Diff Norm	Prod.: Con-Diff High
ID: 2241241+	ID: 2241242+	ID: 2241243+
Exp.: 5-May-2025	Exp.: 5-May-2025	Exp.: 7-May-2025

1:  B o u l e C B A V 2	 B o u l e C B A V 2	 B o u l e C B A V 2
2:  2241241+	 2241242+	 2241243+
3:  25050500001153030000	 25050500002153030000	 25050700003153030000
4:  10004030200000000000	 10004029500000000000	 10004029500000000000
5:  22612648500550314625	 42218833501240435230	 50122921501610546135
6:  08215113152431537730	 24230116152941535230	 51460118153211534930
7:  03403015040020301708	 08806041060070504010	 20818126150131006925
8:  14640428700635050970	 13040466500765045850	 13040605500625033350
9:  11119648825185134763	 11112920886147912503	 11110143870492692946

Boule Con-Diff

CONTROL

IVD



Control	Lot	
Tri-Level	--	--
Low	22412-41	2025-05-05
Normal	22412-42	2025-05-05
High	22412-43	2025-05-07

Open vial stability 14 days ¹

SWELAB

Swelab Alfa/Alfa Plus	CONTROL LOT	L 22412-41	
		Mean ²	± Range ³
RBC	10 ¹² /L	2.28	0.12
MCV	fL	66.1	5.0
RDW	%	14.0	4.0
Hct	%	15.1	2.5
Plt	10 ⁹ /L	82	15
MPV	fL	12.1	1.5
WBC	10 ⁹ /L	3.4	0.3
Hgb	g/dL	5.6	0.3
MCH	pg	24.6	1.5
MCHC	g/dL	37.1	3.0
Lymph#	10 ⁹ /L	1.6	0.4
Mid#	10 ⁹ /L	0.1	0.3
Gran#	10 ⁹ /L	1.7	0.8
Lymph	%	46.9	7.0
Mid	%	3.9	5.0
Gran	%	49.2	7.0

CONTROL LOT	N 22412-42	
	Mean ²	± Range ³
	4.23	0.18
	84.2	5.0
	12.1	4.0
	35.6	3.0
	247	30
	12.5	1.5
	8.7	0.6
	12.5	0.4
	29.6	1.5
	35.1	3.0
	4.5	0.6
	0.4	0.5
	3.8	1.0
	51.3	5.0
	4.5	5.0
	44.2	5.0

CONTROL LOT	H 22412-43	
	Mean ²	± Range ³
	4.99	0.22
	93.4	5.0
	12.1	4.0
	46.6	3.5
	527	60
	12.7	1.5
	20.6	1.8
	16.0	0.5
	32.1	1.5
	34.3	3.0
	13.1	1.5
	0.7	1.0
	6.8	2.5
	63.6	5.0
	3.6	5.0
	32.8	5.0

- 1
- bg Трайност на отворен флакон 14 дни (bg)
 - hr Stabilitnost otvorene bočice 14 dani (hr)
 - cs Stabilita otevřené ampulky 14 dní (cs)
 - da Åben rør stabilitet 14 dage (da)
 - et Stabiilsus avatud vialis 14 päevade (et)
 - fr Stabilité en flacon ouvert 14 jour (fr)
 - de Stabilität geöffneter Flaschen 14 tage (de)
 - el Διατήρηση δείγματος μετά την αποσφράγιση 14 ημέρα (el)
 - it Stabilità della fiala aperta 14 giorni (it)
 - lv Atvērt flakonu stabilitāti 14 dienas (lv)
 - lt Stabilumas atidarius buteliuką 14 dienas (lt)
 - no Åpen rør/glass stabilitet 14 dager (no)
 - pl Trwałość po otwarciu fiołki 14 Liczba dni (pl)
 - pt Estabilidade após abertura do frasco 14 dias (pt)
 - ro Termenul de valabilitate al fiolei desfăcute 14 zile (ro)
 - ru Стабильность вскрытого флакона 14 дней (ru)
 - sr Proizvod stabilan nakon otvaranja pakovanja 14 dana (sr)
 - sk Stabilita otvorenej injekčnej liekovky 14 dní (sk)
 - sl Stabilita otvorenej injekčnej liekovky 14 dni (sl)
 - es Estabilidad de la cápsula abierta 14 días (es)
 - sv Hållbarhet för öppnad flaska 14 dagar (sv)
 - tr Açık şişe stabilitesi 14 gün (tr)

- 2
- Средно
 - Sredstvo
 - Průměr
 - Middelværdi
 - Keskmine
 - Moyenne
 - Mittelwert
 - Μέση
 - Media
 - Nozīmē
 - Vidurkis
 - Middelverdi
 - Średnia
 - Média
 - Medie
 - Среднее
 - Srednja vrednost
 - Priemer
 - Povprečno
 - Media
 - Medelvärde
 - Ortalama

- 3
- Обхват
 - Raspon
 - Rozsah
 - Område
 - Vahemik
 - Intervalle
 - Bereich
 - Αναμενόμενο εύρος α
 - Intervallo
 - Diapazons
 - Intervalas
 - Område
 - Zakres
 - Intervalo
 - Interval
 - диапазон
 - Orseg
 - Rozsah
 - Območje
 - Intervall
 - Aralığı

For Swelab Alfa / Alfa Plus with barcode reader

Alfa:

1. In the Main Menu go to "Q/C Menu" "Enter Con/Cal" "Control/Calibrator Assay Value Input"
2. Scan in each blood calibrator/control column by column, sequence 1 to 9

Alfa Plus:

1. In the Main Menu go to "Quality Control" "Input Assays"
2. Scan in each blood calibrator / control column by column, sequence 1 to 9

Instr.: Swelab Alfa
 Lot: 22412-41
 Prod.: Con-Diff Low
 ID: 2241241+
 Exp.: 5-May-2025

Instr.: Swelab Alfa
 Lot: 22412-42
 Prod.: Con-Diff Norm
 ID: 2241242+
 Exp.: 5-May-2025

Instr.: Swelab Alfa
 Lot: 22412-43
 Prod.: Con-Diff High
 ID: 2241243+
 Exp.: 7-May-2025

