

# Boule Con-Diff

CONTROL

IVD



Control	Lot	
Tri-Level	22502-1K	2025-07-08
Low	22502-11	2025-07-08
Normal	22502-12	2025-07-08
High	22502-13	2025-07-08

Open vial stability 14 days <sup>1</sup>

MEDONIC		CONTROL	L
Medonic M-series/M32		LOT	22502-11
		Mean <sup>2</sup>	± Range <sup>3</sup>
RBC	10 <sup>12</sup> /L	2.26	0.12
MCV	fL	65.1	5.0
RDW	%	15.8	4.0
Hct	%	14.7	2.5
Plt	10 <sup>9</sup> /L	87	15
MPV	fL	11.1	1.5
WBC	10 <sup>9</sup> /L	3.5	0.3
Hgb	g/dL	5.4	0.3
MCH	pg	23.9	1.5
MCHC	g/dL	36.7	3.0
Lymph#	10 <sup>9</sup> /L	1.5	0.4
Mid#	10 <sup>9</sup> /L	0.2	0.3
Gran#	10 <sup>9</sup> /L	1.8	0.8
Lymph	%	43.1	7.0
Mid	%	5.9	5.0
Gran	%	51.0	7.0

CONTROL	N	
LOT	22502-12	
	Mean <sup>2</sup>	± Range <sup>3</sup>
	4.18	0.18
	83.8	5.0
	12.9	4.0
	35.0	3.0
	239	30
	11.4	1.5
	8.5	0.6
	12.3	0.4
	29.4	1.5
	35.1	3.0
	4.0	0.6
	0.6	0.5
	3.9	1.0
	46.8	5.0
	6.8	5.0
	46.4	5.0

CONTROL	H	
LOT	22502-13	
	Mean <sup>2</sup>	± Range <sup>3</sup>
	5.11	0.22
	94.1	5.0
	14.1	4.0
	48.1	3.5
	535	60
	11.6	1.5
	20.4	1.8
	16.5	0.5
	32.3	1.5
	34.3	3.0
	12.5	1.5
	1.2	1.0
	6.7	2.5
	61.4	5.0
	5.8	5.0
	32.8	5.0

- 1
- bg Трайност на отворен флакон 14 дни (bg)
  - hr Stabilnost otvorene bočice 14 dani (hr)
  - cs Stabilita otevřené ampulky 14 dní (cs)
  - da Åben rør stabilitet 14 dage (da)
  - et Stabiilsus avatud vialis 14 päevade (et)
  - fr Stabilité en flacon ouvert 14 jour (fr)
  - de Stabilität geöffneter Flaschen 14 tage (de)
  - el Διατήρηση δείγματος μετά την αποσφράγιση 14 ημέρα (el)
  - it Stabilità della fiala aperta 14 giorni (it)
  - lv Atvērt flakonu stabilitāti 14 dienas (lv)
  - lt Stabilumas atidarius buteliuką 14 dienas (lt)
  - no Åpen rør/glass stabilitet 14 dager (no)
  - pl Trwałość po otwarciu fiołki 14 Liczba dni (pl)
  - pt Estabilidade após abertura do frasco 14 dias (pt)
  - ro Termenul de valabilitate al fiolei desfăcute 14 zile (ro)
  - ru Стабильность вскрытого флакона 14 дней (ru)
  - sr Proizvod stabilan nakon otvaranja pakovanja 14 dana (sr)
  - sk Stabilita otvorenej injekčnej liekovky 14 dní (sk)
  - sl Stabilita otvorenej injekčnej liekovky 14 dni (sl)
  - es Estabilidad de la cápsula abierta 14 días (es)
  - sv Hållbarhet för öppnad flaska 14 dagar (sv)
  - tr Açık şişe stabilitesi 14 gün (tr)

- 2
- Средно
  - Sredstvo
  - Průměr
  - Middelværdi
  - Keskmine
  - Moyenne
  - Mittelwert
  - Μέση
  - Media
  - Nozīmē
  - Vidurkis
  - Middelverdi
  - Średnia
  - Média
  - Medie
  - Среднее
  - Srednja vrednost
  - Priemer
  - Povprečno
  - Media
  - Medelvärde
  - Ortalama

- 3
- Обхват
  - Raspon
  - Rozsah
  - Område
  - Vahemik
  - Intervalle
  - Bereich
  - Αναμενόμενο εύρος α
  - Intervallo
  - Diapazons
  - Intervalas
  - Område
  - Zakres
  - Intervalo
  - Interval
  - диапазон
  - Opseg
  - Rozsah
  - Območje
  - Intervalo
  - Intervall
  - Aralığı

# For Medonic M-series / M32 with barcode reader

**M-Series:**

1. In the Main Menu go to "Q/C Menu" "Enter Con/Cal" "Control/Calibrator Assay Value Input"
2. Scan in each blood calibrator/control column by column, sequence 1 to 9

**M32:**

1. In the Main Menu go to "Quality Control" "Input Assays"
2. Scan in each blood calibrator / control column by column, sequence 1 to 9

Instr.: M-series  
 Lot: 22502-11  
 Prod.: Con-Diff Low  
 ID: 2250211+  
 Exp.: 8-Jul-2025

Instr.: M-series  
 Lot: 22502-12  
 Prod.: Con-Diff Norm  
 ID: 2250212+  
 Exp.: 8-Jul-2025

Instr.: M-series  
 Lot: 22502-13  
 Prod.: Con-Diff High  
 ID: 2250213+  
 Exp.: 8-Jul-2025



3.

# Boule Con-Diff

CONTROL

IVD



Control	Lot	
Tri-Level	22502-1K	2025-07-08
Low	22502-11	2025-07-08
Normal	22502-12	2025-07-08
High	22502-13	2025-07-08

Open vial stability 14 days <sup>1</sup>

## SWELAB

CONTROL		L	
LOT		22502-11	
Swelab Alfa/Alfa Plus		Mean <sup>2</sup>	± Range <sup>3</sup>
RBC	10 <sup>12</sup> /L	2.28	0.12
MCV	fL	66.2	5.0
RDW	%	15.2	4.0
Hct	%	15.1	2.5
Plt	10 <sup>9</sup> /L	88	15
MPV	fL	11.8	1.5
WBC	10 <sup>9</sup> /L	3.5	0.3
Hgb	g/dL	5.5	0.3
MCH	pg	24.1	1.5
MCHC	g/dL	36.4	3.0
Lymph#	10 <sup>9</sup> /L	1.6	0.4
Mid#	10 <sup>9</sup> /L	0.1	0.3
Gran#	10 <sup>9</sup> /L	1.8	0.8
Lymph	%	46.8	7.0
Mid	%	3.2	5.0
Gran	%	50.0	7.0

CONTROL		N	
LOT		22502-12	
		Mean <sup>2</sup>	± Range <sup>3</sup>
		4.22	0.18
		84.8	5.0
		11.9	4.0
		35.8	3.0
		245	30
		12.1	1.5
		8.6	0.6
		12.4	0.4
		29.4	1.5
		34.6	3.0
		4.4	0.6
		0.3	0.5
		3.9	1.0
		50.5	5.0
		3.7	5.0
		45.8	5.0

CONTROL		H	
LOT		22502-13	
		Mean <sup>2</sup>	± Range <sup>3</sup>
		5.13	0.22
		95.3	5.0
		13.0	4.0
		48.9	3.5
		557	60
		12.5	1.5
		20.4	1.8
		16.5	0.5
		32.2	1.5
		33.7	3.0
		13.2	1.5
		0.7	1.0
		6.5	2.5
		64.7	5.0
		3.4	5.0
		31.9	5.0

1

bg	Трайност на отворен флакон 14 дни (bg)
hr	Stabilnost otvorene bočice 14 dani (hr)
cs	Stabilita otevřené ampulky 14 dní (cs)
da	Åben rør stabilitet 14 dage (da)
et	Stabiilsus avatud viaalis 14 päevade (et)
fr	Stabilité en flacon ouvert 14 jour (fr)
de	Stabilität geöffneter Flaschen 14 tage (de)
el	Διατήρηση δείγματος μετά την αποσφράγιση 14 ημέρα (el)
it	Stabilità della fiala aperta 14 giorni (it)
lv	Atvērt flakonu stabilitāti 14 dienas (lv)
lt	Stabilumas atidarius buteliuką 14 dienas (lt)
no	Åpen rør/glass stabilitet 14 dager (no)
pl	Trwałość po otwarciu fiołki 14 Liczba dni (pl)
pt	Estabilidade após abertura do frasco 14 dias (pt)
ro	Termenul de valabilitate al fiolei desfăcute 14 zile (ro)
ru	Стабильность вскрытого флакона 14 дней (ru)
sr	Proizvod stabilan nakon otvaranja pakovanja 14 dana (sr)
sk	Stabilita otvorenej injekčnej liekovky 14 dní (sk)
sl	Stabilita otvorenej injekčnej liekovky 14 dni (sl)
es	Estabilidad de la cápsula abierta 14 días (es)
sv	Hållbarhet för öppnad flaska 14 dagar (sv)
tr	Açık şişe stabilitesi 14 gün (tr)

2

	Средно
	Sredstvo
	Průměr
	Middelverdidi
	Keskmine
	Moyenne
	Mittelwert
	Μέση
	Media
	Nozīmē
	Vidurkis
	Middelverdi
	Średnia
	Média
	Medie
	Среднее
	Srednja vrednost
	Priemer
	Povprečno
	Media
	Medelvärde
	Ortalama

3

	Обхват
	Raspon
	Rozsah
	Område
	Vahemik
	Intervalle
	Bereich
	Αναμενόμενο εύρος α
	Intervallo
	Diapazons
	Intervalas
	Område
	Zakres
	Intervalo
	Interval
	диапазон
	Orseg
	Rozsah
	Območje
	Intervalo
	Intervall
	Aralığı

## For Swelab Alfa / Alfa Plus with barcode reader

### Alfa:

1. In the Main Menu go to "Q/C Menu" "Enter Con/Cal" "Control/Calibrator Assay Value Input"
2. Scan in each blood calibrator/control column by column, sequence 1 to 9

### Alfa Plus:

1. In the Main Menu go to "Quality Control" "Input Assays"
2. Scan in each blood calibrator / control column by column, sequence 1 to 9

Instr.: Swelab Alfa  
Lot: 22502-11  
Prod.: Con-Diff Low  
ID: 2250211+  
Exp.: 8-Jul-2025

Instr.: Swelab Alfa  
Lot: 22502-12  
Prod.: Con-Diff Norm  
ID: 2250212+  
Exp.: 8-Jul-2025

Instr.: Swelab Alfa  
Lot: 22502-13  
Prod.: Con-Diff High  
ID: 2250213+  
Exp.: 8-Jul-2025

