

Boule Con-Diff

CONTROL

IVD



Open vial stability 14 days ¹

Control	Lot	
Tri-Level	--	--
Low	22502-41	2025-07-08
Normal	22502-42	2025-07-08
High	22502-43	2025-07-08

MEDONIC

Medonic M-series/M32		CONTROL	
		LOT	L 22502-41
		Mean ²	± Range ³
RBC	10 ¹² /L	2.26	0.12
MCV	fL	65.1	5.0
RDW	%	15.8	4.0
Hct	%	14.7	2.5
Plt	10 ⁹ /L	87	15
MPV	fL	11.1	1.5
WBC	10 ⁹ /L	3.5	0.3
Hgb	g/dL	5.4	0.3
MCH	pg	23.9	1.5
MCHC	g/dL	36.7	3.0
Lymph#	10 ⁹ /L	1.5	0.4
Mid#	10 ⁹ /L	0.2	0.3
Gran#	10 ⁹ /L	1.8	0.8
Lymph	%	43.1	7.0
Mid	%	5.9	5.0
Gran	%	51.0	7.0

	CONTROL	
	LOT	N 22502-42
	Mean ²	± Range ³
	4.18	0.18
	83.8	5.0
	12.9	4.0
	35.0	3.0
	239	30
	11.4	1.5
	8.5	0.6
	12.3	0.4
	29.4	1.5
	35.1	3.0
	4.0	0.6
	0.6	0.5
	3.9	1.0
	46.8	5.0
	6.8	5.0
	46.4	5.0

	CONTROL	
	LOT	H 22502-43
	Mean ²	± Range ³
	5.11	0.22
	94.1	5.0
	14.1	4.0
	48.1	3.5
	535	60
	11.6	1.5
	20.4	1.8
	16.5	0.5
	32.3	1.5
	34.3	3.0
	12.5	1.5
	1.2	1.0
	6.7	2.5
	61.4	5.0
	5.8	5.0
	32.8	5.0

1

bg	Трайност на отворен флакон 14 дни (bg)
hr	Stabilitnost otvorene bočice 14 dani (hr)
cs	Stabilita otevřené ampulky 14 dní (cs)
da	Åben rør stabilitet 14 dage (da)
et	Stabiilsus avatud vialis 14 päevade (et)
fr	Stabilité en flacon ouvert 14 jour (fr)
de	Stabilität geöffneter Flaschen 14 tage (de)
el	Διατήρηση δείγματος μετά την αποσφράγιση 14 ημέρα (el)
it	Stabilità della fiala aperta 14 giorni (it)
lv	Atvērt flakonu stabilitāti 14 dienas (lv)
lt	Stabilumas atidarius buteliuką 14 dienas (lt)
no	Åpen rør/glass stabilitet 14 dager (no)
pl	Trwałość po otwarciu fiołki 14 Liczba dni (pl)
pt	Estabilidade após abertura do frasco 14 dias (pt)
ro	Termenul de valabilitate al fiolei desfăcute 14 zile (ro)
ru	Стабильность вскрытого флакона 14 дней (ru)
sr	Proizvod stabilan nakon otvaranja pakovanja 14 dana (sr)
sk	Stabilita otvorenej injekčnej liekovky 14 dní (sk)
sl	Stabilita otvorenej injekčnej liekovky 14 dni (sl)
es	Estabilidad de la cápsula abierta 14 días (es)
sv	Hållbarhet för öppnad flaska 14 dagar (sv)
tr	Açık şişe stabilitesi 14 gün (tr)

2

	Средно
	Sredstvo
	Průměr
	Middelværdi
	Keskmine
	Moyenne
	Mittelwert
	Μέση
	Media
	Nozīmē
	Vidurkis
	Middelverdi
	Średnia
	Média
	Medie
	Среднее
	Srednja vrednost
	Priemer
	Povprečno
	Media
	Medelvärde
	Ortalama

3

	Обхват
	Raspon
	Rozsah
	Område
	Vahemik
	Intervalle
	Bereich
	Αναμενόμενο εύρος α
	Intervallo
	Diapazons
	Intervalas
	Område
	Zakres
	Intervalo
	Interval
	диапазон
	Opseg
	Rozsah
	Območje
	Intervall
	Intervall
	Aralığı

For Medonic M-series / M32 with barcode reader




M-Series:

1. In the Main Menu go to "Q/C Menu" "Enter Con/Cal" "Control/Calibrator Assay Value Input"
2. Scan in each blood calibrator/control column by column, sequence 1 to 9

M32:

1. In the Main Menu go to "Quality Control" "Input Assays"
2. Scan in each blood calibrator / control column by column, sequence 1 to 9

Instr.: M-series	Instr.: M-series	Instr.: M-series
Lot: 22502-41	Lot: 22502-42	Lot: 22502-43
Prod.: Con-Diff Low	Prod.: Con-Diff Norm	Prod.: Con-Diff High
ID: 2250241+	ID: 2250242+	ID: 2250243+
Exp.: 8-Jul-2025	Exp.: 8-Jul-2025	Exp.: 8-Jul-2025

1:  B o u l e C B A V 2	 B o u l e C B A V 2	 B o u l e C B A V 2
2:  2250241+	 2250242+	 2250243+
3:  25070800001153030000	 25070800002153030000	 25070800003153030000
4:  10004030000000000000	 10004029700000000000	 10004029700000000000
5:  22612651500540314725	 41818838501230435030	 51122941501650548135
6:  08715111152391536730	 23930114152941535130	 53560116153231534330
7:  03503015040020301808	 08506040060060503910	 20418125150121006725
8:  15840431700595051070	 12940468500685046450	 14140614500585032850
9:  11116208343503311759	 11110953176963261252	 11110678530530725479

Boule Con-Diff

CONTROL

IVD



Control	Lot	
Tri-Level	--	--
Low	22502-41	2025-07-08
Normal	22502-42	2025-07-08
High	22502-43	2025-07-08

Open vial stability 14 days ¹

SWELAB

Swelab Alfa/Alfa Plus	CONTROL		L
	LOT	22502-41	22502-41
	Mean ²	± Range ³	
RBC	10 ¹² /L	2.28	0.12
MCV	fL	66.2	5.0
RDW	%	15.2	4.0
Hct	%	15.1	2.5
Plt	10 ⁹ /L	88	15
MPV	fL	11.8	1.5
WBC	10 ⁹ /L	3.5	0.3
Hgb	g/dL	5.5	0.3
MCH	pg	24.1	1.5
MCHC	g/dL	36.4	3.0
Lymph#	10 ⁹ /L	1.6	0.4
Mid#	10 ⁹ /L	0.1	0.3
Gran#	10 ⁹ /L	1.8	0.8
Lymph	%	46.8	7.0
Mid	%	3.2	5.0
Gran	%	50.0	7.0

CONTROL	N	
	LOT	22502-42
	Mean ²	± Range ³
	4.22	0.18
	84.8	5.0
	11.9	4.0
	35.8	3.0
	245	30
	12.1	1.5
	8.6	0.6
	12.4	0.4
	29.4	1.5
	34.6	3.0
	4.4	0.6
	0.3	0.5
	3.9	1.0
	50.5	5.0
	3.7	5.0
	45.8	5.0

CONTROL	H	
	LOT	22502-43
	Mean ²	± Range ³
	5.13	0.22
	95.3	5.0
	13.0	4.0
	48.9	3.5
	557	60
	12.5	1.5
	20.4	1.8
	16.5	0.5
	32.2	1.5
	33.7	3.0
	13.2	1.5
	0.7	1.0
	6.5	2.5
	64.7	5.0
	3.4	5.0
	31.9	5.0

- 1
- bg Трайност на отворен флакон 14 дни (bg)
 - hr Stabilitnost otvorene bočice 14 dani (hr)
 - cs Stabilita otevrené ampulky 14 dny (cs)
 - da Åben rør stabilitet 14 dage (da)
 - et Stabiilsus avatud viaalis 14 päevade (et)
 - fr Stabilité en flacon ouvert 14 jour (fr)
 - de Stabilität geöffneter Flaschen 14 tage (de)
 - el Διατήρηση δείγματος μετά την αποσφράγιση 14 ημέρα (el)
 - it Stabilità della fiala aperta 14 giorni (it)
 - lv Atvērt flakonu stabilitāti 14 dienas (lv)
 - lt Stabilumas atidarius buteliuką 14 dienas (lt)
 - no Åpen rør/glass stabilitet 14 dager (no)
 - pl Trwałość po otwarciu fiołki 14 Liczba dni (pl)
 - pt Estabilidade após abertura do frasco 14 dias (pt)
 - ro Termenul de valabilitate al fiolei desfacute 14 zile (ro)
 - ru Стабильность вскрытого флакона 14 дней (ru)
 - sr Proizvod stabilan nakon otvaranja pakovanja 14 dana (sr)
 - sk Stabilita otvorenej injekčnej liekovky 14 dní (sk)
 - sl Stabilita otvorenej injekčnej liekovky 14 dni (sl)
 - es Estabilidad de la cápsula abierta 14 días (es)
 - sv Hållbarhet för öppnad flaska 14 dagar (sv)
 - tr Açık şişe stabilitesi 14 gün (tr)

- 2
- Средно
 - Sredstvo
 - Průměr
 - Middelverdidi
 - Keskmine
 - Moyenne
 - Mittelwert
 - Μέση
 - Media
 - Nozīmē
 - Vidurkis
 - Middelverdi
 - Średnia
 - Média
 - Medie
 - Среднее
 - Srednja vrednost
 - Priemer
 - Povprečno
 - Media
 - Medelvärde
 - Ortalama

- 3
- Обхват
 - Raspon
 - Rozsah
 - Område
 - Vahemik
 - Intervalle
 - Bereich
 - Αναμενόμενο εύρος α
 - Intervallo
 - Diapazons
 - Intervalas
 - Område
 - Zakres
 - Intervalo
 - Interval
 - диапазон
 - Orseg
 - Rozsah
 - Območje
 - Intervalo
 - Intervall
 - Aralığı

For Swelab Alfa / Alfa Plus with barcode reader

Alfa:

1. In the Main Menu go to "Q/C Menu" "Enter Con/Cal" "Control/Calibrator Assay Value Input"
2. Scan in each blood calibrator/control column by column, sequence 1 to 9

Alfa Plus:

1. In the Main Menu go to "Quality Control" "Input Assays"
2. Scan in each blood calibrator / control column by column, sequence 1 to 9

Instr.: Swelab Alfa
Lot: 22502-41
Prod.: Con-Diff Low
ID: 2250241+
Exp.: 8-Jul-2025

Instr.: Swelab Alfa
Lot: 22502-42
Prod.: Con-Diff Norm
ID: 2250242+
Exp.: 8-Jul-2025

Instr.: Swelab Alfa
Lot: 22502-43
Prod.: Con-Diff High
ID: 2250243+
Exp.: 8-Jul-2025

