

## CERTIFICATE OF ANALYSIS

**Product:** Swelab™Alfalyse, 5L

**Lot No:** 2406-004

**Product Part #:** 501-207

**Exp. Date:** 2027-05-31

**Manufacturing Date:** 2024-06-17

**Quality Testing Date:** 2024-06-20

Parameter	Specification	Result	Pass/Fail
pH (units):	7.54 – 8.30	8.17	Pass
Osmolality (mOsm/kg):	295 – 330	317	Pass
Conductivity (mS/cm):	18.00– 18.50	18.16	Pass
Functional Test	Pass	Pass	Pass

The product has been manufactured in compliance with Quality Systems Regulations and has met all current release specifications.

*Alexander*

Quality Assurance

*06/21/24*

Date

