

# Certificate of Analysis

<b>Product</b>	SWELAB LUMI-L2 LYSE (500 mL)
<b>Product No</b>	1504516
<b>Lot No</b>	2024061401
<b>Manufacturing Date</b>	2024-06-14
<b>Expiration Date</b>	2026-06-13

<b>Test</b>	<b>Specification</b>
pH	4,0 – 4,4
CBC Performance test*	Passes test
Background WBC (x 10 <sup>9</sup> /L)	≤ 0,2

\*Performance testing at a reference laboratory. WBC, HGB and WBC scattergrams tested and approved.

Actual values meet specification.

Date: 2024-10-30



Ulrika Miefalk  
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.  
All information in this document is property of Boule Medical AB.

