


# Certificate of Analysis

**Product** SWELAB LUMI-D DILUENT (20L)  
**Product No** 1504514  
**Lot No** 2024062002  
**Manufacturing Date** 2024-06-20  
**Expiration Date** 2026-06-19

Test	Specification
pH	7,1 – 7,5
Osmolality (mOsm/kg)	290 – 310
Conductivity (mS/cm)	17,3 – 18,3
Particles PLT/RBC channel (x 10 <sup>9</sup> /L)	≤ 5

Actual values meet specification.

Date: 2024-09-11



*Caroline Sparf*  
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.  
All information in this document is property of Boule Medical AB.

