

Certificate of Analysis

Product SWELAB LUMI-L2 LYSE (500 mL)
Product No 1504516
Lot No 2024111301
Manufacturing Date 2024-11-13
Expiration Date 2026-11-12

Test	Specification
pH	4,0 – 4,4
CBC Performance test*	Passes test
Background WBC ($\times 10^9/L$)	$\leq 0,2$

*Performance testing at a reference laboratory. WBC, HGB and WBC scattergrams tested and approved.

Actual values meet specification.

Date: 2025-02-06



Caroline Sparf
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

