

Certificate of Analysis

Product SWELAB LUMI-L2 LYSE (500 mL)
Product No 1504516
Lot No 2025010901
Manufacturing Date 2025-01-09
Expiration Date 2027-01-08

| Test | Specification |
|---------------------------------------|---------------|
| pH | 4,0 – 4,4 |
| CBC Performance test* | Passes test |
| Background WBC (x 10 ⁹ /L) | ≤ 0,2 |

*Performance testing at a reference laboratory. WBC, HGB and WBC scattergrams tested and approved.

Actual values meet specification.

Date: 2025-04-07



Ulrika Miefalk
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

