

Certificate of Analysis

Product H50V SHEATH, RFID
Product No 1440107 (5 L)
Lot No 24011384M5
Manufacturing Date 2024-03-15
Expiration Date 2026-03-14

Test	Specification
Appearance	Sheath should be transparent liquid without precipitation, particles or floccules
pH	8.10 - 8.70
Osmolality (mOsm/kg)	20 - 60
Blanks	WBC $\leq 0.2 \times 10^9/L$

Actual values meet specifications.

Date: 2024-04-23



Sofia Karlström
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

