

# Certificate of Analysis

**Product** H50V DILUENT, RFID  
**Product No** 1440103 (5 L)  
**Lot No** 24020041M5  
**Manufacturing Date** 2024-03-13  
**Expiration Date** 2026-03-12

Test	Specification
Appearance	Diluent should be transparent liquid without precipitation, particle or floccules
pH	6.80 – 7.00
Osmolality (mOsm/kg)	309 – 321
Conductivity (mS/cm)	18.50 - 19.50
Particles ( $10^5/L$ ) (Particle volume $\geq$ 2.5fL)	$\leq 2.5$

Actual values meet specifications.

Date: 2024-04-23



*Sofia Karlström*  
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.  
All information in this document is property of Boule Medical AB.

