

# Certificate of Analysis

**Product** Quintus 5-part Lyse 5L  
**Product No** 1504292  
**Lot No** 2404-178  
**Manufacturing Date** 2024-04-17  
**Expiration Date** 2025-10-16

Test	Specification
pH	2,7 – 3,0
Osmolality (mOsm/kg)	35 – 44
Conductivity ( $\mu\text{S}/\text{cm}$ )	1117 – 1541
Background WBC ( $\times 10^{12}/\text{L}$ )	$\leq 0.1$

Actual values meet specification.

Date: 2024-06-04



*Pernilla Hammarstedt Rydberg*  
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.  
All information in this document is property of Boule Medical AB.

