

# Certificate of Analysis

**Product** BOULE ENZYMATIC CLEANER (100 mL)  
**Product No** 1504112  
**Lot No** 2404-180  
**Manufacturing Date** 2024-04-18  
**Expiration Date** 2026-04-17

Test Enzymatic Cleaner	Specification
pH	7.6– 8.4
Blank	PLT < 100 x 10 <sup>9</sup> /L RBC < 0.05 x 10 <sup>12</sup> /L

Actual values meet specifications.

Date: 2024-04-26

  
Ulrika Miefalk  
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.  
All information in this document is property of Boule Medical AB.

