

# Certificate of Analysis

**Product** H50V SHEATH, RFID  
**Product No** 1440108 (1.9 L)  
**Lot No** 24041592M6  
**Manufacturing Date** 2024-05-07  
**Expiration Date** 2026-05-06

Test	Specification
Appearance	Sheath should be transparent liquid without precipitation, particles or floccules
pH	8.10 - 8.70
Osmolality (mOsm/kg)	20 - 60
Blanks	WBC $\leq 0.2 \times 10^9/L$

Actual values meet specifications.

Date: 2024-08-21



*Pernilla Hammarstedt Rydberg*  
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.  
All information in this document is property of Boule Medical AB.

