

Certificate of Analysis

| | |
|---------------------------|-----------------------------|
| Product | SWELAB ALFALYSE, RFID (5 L) |
| Product No | 1504463 |
| Lot No | 2406-207 |
| Manufacturing Date | 2024-06-17 |
| Expiration Date | 2027-06-16 |

| Test | Specification |
|-------------------------|----------------------|
| pH | 7,80 – 8,30 |
| Osmolality (mOsm/kg) | 295 – 330 |
| Conductivity (mS/cm) | 17,75 – 18,25 |
| CBC Performance * | Passes test |
| Particles (WBC channel) | ≤ 0,2 |

* Performance test using calibrated Blood cell counter at Boule Medical reference laboratory. All parameters except mid cells are tested. Accepted reference 95 % to 105 % of expected control blood sample values.

Actual value meets specification.

Date: 2024-06-25



Caroline Sparf
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

