


# Certificate of Analysis

**Product** H50V Probe Cleaner  
**Product No** 1440111 (100 mL)  
**Lot No** 24081373M2  
**Manufacturing Date** 2024-09-06  
**Expiration Date** 2025-09-05

Test	Specification
Appearance	Transparent liquid without precipitation, particle or floccules
pH	12,00 - 14,00
Blank	WBC $\leq 0.5 \times 10^9/L$ RBC $\leq 0.1 \times 10^{12}/L$ HGB $\leq 5g/L$ PLT $\leq 20 \times 10^9/L$

Actual values meet specifications.

Date: 2024-09-17



Caroline Sparf  
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.  
All information in this document is property of Boule Medical AB.

