

Certificate of Analysis

Product	Boule Micropipette Plastic EDTA 10x100 pcs Boule Micropipette Plastic EDTA 1x100 pcs
Product No	1070030 1070039
Lot No	240845K7
Manufacturing Date	2024-04-11
Expiration Date	2027-03-28

Parameter	Specification
Volume 20 μ L	$CV \leq 1,0\%$
WBC background*	$\leq 0,1 \times 10^9/L$
RBC background*	$\leq 0,01 \times 10^{12}/L$
HGB background	$\leq 0,2 \text{ g/dL}$
PLT background	$\leq 10 \times 10^9/L$
*WBC background for Exigo	$\leq 0,2 \times 10^9/L$
*RBC background for Exigo	$\leq 0,03 \times 10^{12}/L$

Actual value meets specification.

Date: 2024-06-05



Ulrika Miefalk
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

