

Certificate of Analysis

Product	Boule Micropipette Plastic EDTA 10x100 pcs Boule Micropipette Plastic EDTA 1x100 pcs
Product No	1070030 1070039
Lot No	242245D6
Manufacturing Date	2024-10-15
Expiration Date	2027-09-28

Parameter	Specification
Volume 20 µL	CV ≤ 1,0%
WBC background*	≤ 0,1 x 10 ⁹ /L
RBC background*	≤ 0,01 x 10 ¹² /L
HGB background	≤ 0,2 g/dL
PLT background	≤ 10 x 10 ⁹ /L
*WBC background for Exigo	≤ 0,2 x 10 ⁹ /L
*RBC background for Exigo	≤ 0,03 x 10 ¹² /L

Actual value meets specification.

Date: 2025-02-14



Caroline Sparf
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

