

CERTIFICATE OF ANALYSIS

Product: Swelab™AlfaDiluent

Lot No: 2501-004

Product Part #: 501-208


Exp. Date: 2027-12-31

Manufacturing Date: 01-17-2025


Quality Testing Date: 01-22-2025

Parameter	Specification	Result	Pass/Fail
pH (units):	6.50 – 6.90	6.78	Pass
Osmolality (mOsm/kg):	294 – 312	297	Pass
Conductivity (mS/cm):	16.98 – 17.93	17.29	Pass
Background (x10 ³ /μL):	0.0 – 8.0	2.4	Pass

The product has been manufactured in compliance with Quality Systems Regulations and has met all current release specifications.



Quality Assurance



Date

