

Certificate of Analysis

Product EXIGO DILUENT
Product No 1504201 (1.9 L)
Lot No 2502-280
Manufacturing Date 2025-02-17
Expiration Date 2028-02-16

Test	Specification
pH	6.90 – 7.20
Osmolality (mOsm/kg)	300 – 330
Conductivity (mS/cm)	18.00 - 19.00
Particles (PLT/RBC channel)	≤ 5

Actual values meet specifications.

Date: 2025-02-20



Elin Andersson
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

