

Certificate of Analysis


Product EXIGO EOS REAGENT (1.9 L)
Product No 1504278
Lot No 2503-287
Manufacturing Date 2025-03-10
Expiration Date 2027-03-09

| Test | Specification |
|----------------------|---------------|
| pH | 8.70 – 9.30 |
| Osmolality (mOsm/kg) | 125 - 145 |
| Conductivity (mS/cm) | 6.80 – 7.60 |
| EOS Performance * | Passes test |
| Blank EOS | ≤ 0.1 |

* Performance test using calibrated Blood cell counter at Boule Medical reference laboratory.

Actual values meet specifications.

Date: 2025-03-10



Ulrika Miefalk
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

