

# Certificate of Analysis

**Product** EXIGO DILUENT  
**Product No** 1504328 (10 L)  
**Lot No** 2504-298  
**Manufacturing Date** 2025-04-01  
**Expiration Date** 2028-03-31

| Test                        | Specification |
|-----------------------------|---------------|
| pH                          | 6.90 – 7.20   |
| Osmolality (mOsm/kg)        | 300 – 330     |
| Conductivity (mS/cm)        | 18.00 - 19.00 |
| Particles (PLT/RBC channel) | ≤ 5           |

Actual values meet specifications.

Date: 2025-04-04



*Pernilla Hammarstedt Rydberg*  
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.  
All information in this document is property of Boule Medical AB.

