

Certificate of Analysis

Product EXIGO EOS REAGENT, RFID (1.9 L)
Product No 1504502
Lot No 2504-304
Manufacturing Date 2025-04-08
Expiration Date 2027-04-07

Test	Specification
pH	8.70 – 9.30
Osmolality (mOsm/kg)	125 - 145
Conductivity (mS/cm)	6.80 – 7.60
EOS Performance *	Passes test
Blank EOS	≤ 0.1

* Performance test using calibrated Blood cell counter at Boule Medical reference laboratory.

Actual values meet specifications.

Date: 2025-04-09



Caroline Sparf
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

