


Certificate of Analysis

Product MEDONIC M-SERIES DILUENT, RFID (20 L)
Product No 1504460
Lot No 2504-306
Manufacturing Date 2025-04-10
Expiration Date 2028-04-09

| Test | Specification |
|-----------------------------|---------------|
| pH | 6,90 – 7,10 |
| Osmolality (mOsm/kg) | 290 – 310 |
| Conductivity (mS/cm) | 17,50 – 18,00 |
| Particles (PLT/RBC channel) | ≤ 5 |

Actual value meets specification.

Date: 2025-04-15



Caroline Sparf
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

