

Boule Con-5Diff A1

CONTROL

IVD

LOT



Tri-Level	72502-1K	2025/05/30
Low	72502-11	2025/05/30
Normal	72502-12	2025/05/30
High	72502-13	2025/05/30

Open vial stability: 14 days ¹

Medonic M51 / Swelab Lumi

Parameter ²	Unit ³	Low		Normal		High	
		Mean ⁴	± Range ⁵	Mean ⁴	± Range ⁵	Mean ⁴	± Range ⁵
WBC	10 ⁹ /L	3.45	± 0.50	7.92	± 1.00	18.02	± 2.50
Neu%	%	48.5	16.0	54.0	8.0	63.1	8.0
Lym%	%	41.7	21.0	32.9	8.0	21.0	12.0
Mon%	%	3.9	5.0	3.7	5.0	4.1	6.0
Eos%	%	5.9	5.0	9.4	6.0	11.8	8.0
Bas%	%	0.3	10.0	0.4	10.0	0.3	10.0
Neu#	10 ⁹ /L	1.68	0.50	4.28	0.70	11.38	2.00
Lym#	10 ⁹ /L	1.44	0.70	2.60	0.80	3.78	2.00
Mon#	10 ⁹ /L	0.14	0.30	0.29	0.50	0.73	1.10
Eos#	10 ⁹ /L	0.20	0.20	0.75	0.60	2.13	1.40
Bas#	10 ⁹ /L	0.01	0.40	0.03	0.80	0.05	1.80
RBC	10 ¹² /L	2.32	0.24	4.66	0.30	5.37	0.50
HGB	g/L	59	6	135	6	171	8
HCT	%	18.8	2.0	42.9	3.0	53.5	4.0
MCV	fL	80.9	5.0	92.0	5.0	99.7	6.0
MCH	pg	25.4	2.5	29.0	2.5	31.8	2.5
MCHC	g/L	314	30	315	30	319	30
RDW-CV	%	15.3	3.0	13.7	3.0	13.1	3.0
PLT	10 ⁹ /L	64	20	262	40	491	60
MPV	fL	8.1	3.0	8.4	3.0	7.8	3.0

No mean values and ranges are provided for research parameters-RDW-SD, PDW, PCT, P-LCR, and P-LCC. User may establish these based on their usage.



<p>1</p> <p>bg Трайност на отворен флакон: 14 дни cs Stabilita otevřené ampulky: 14 dny da Åben rør stabilitet: 14 dage de Stabilität geöffneter Flaschen: 14 tage el Διατήρηση δείγματος μετά την αποσφράγιση: 14 ημέρα es Estabilidad de la cápsula abierta: 14 días et Stabiilsus avatud vialis: 14 päevade arv fr Stabilité en flacon ouvert: 14 jour hr Stabilnost otvorene bočice: 14 dani it Stabilità della fiala aperta: 14 giorni lt Stabilumas atidarius buteliukai: 14 dienų lv Atvērt flakonu stabilitāte: 14 dienas no Åpen rør/ glass stabilitet: 14 dager pl Trwałość po otworzeniu fiolki: 14 Liczba dni pt Estabilidade após abertura do frasco: 14 dias ro Termenul de valabilitate al fiolei desfăcute: 14 zile ru Стабильность вскрытого флакона: 14 дней sk Stabilita otvorenej injekčnej liekovky: 14 dní sl Stabilitost odprte stekleničke: 14 dni sr Proizvod stabilan nakon otvaranja pakovanja: 14 dana sv Hållbarhet för öppnad flaska: 14 dagar tr Acık sise stabilitesi gün: 14 gün</p>	<p>2</p> <p>Параметър Parametr Parameter Parameter Παράμετρος Parámetro Parameeter Paramètre Parametar Parametro Parametras Parametrs Parameter Kryterium Parâmetro Parametru Параметр Parameter Parameter Parametar Parameter Parametre</p>	<p>3</p> <p>Единица Jednotka Enhed Einheit Μονάδα Unidad Ühik Unité Jedinica Unită Vienetas Vienība Enhet Jednostka Unidade Unitate Единица измерения Jednotka Enota Jedinica Enhet Birim</p>	<p>4</p> <p>Средно Průměr Middelværdi Mittelwert Μέση Media Keskmine Moyenne Sredstvo Media Vidurkis Nozīmē Middelverdi Średnia Média Médie Среднее Priemer Povprečno Srednja vrednost Medelvärde Ortalama</p>	<p>5</p> <p>Обхват Rozsah Område Bereich Αναμενόμενο εύρος α Intervalo Vaheмик Intervalle Raspon Intervallo Intervalas Diapazons Område Zakres Intervalo Intervall Intervall Rozsah Območje Opseg Intervall Aralıđı</p>
---	--	---	--	---

1504546 31340-12