

Certificate of Analysis


Product MEDONIC M51-L1 LYSE (200 mL)
Product No 1504511
Lot No 2025020701
Manufacturing Date 2025-02-07
Expiration Date 2027-02-06

Test	Specification
pH	2,0 – 2,4
CBC Performance test*	Passes test
Background WBC ($\times 10^9/L$)	$\leq 0,2$

*Performance testing at a reference laboratory. WBC and HGB tested and approved.

Actual values meet specification.

Date: 2025-05-19



Caroline Sparf
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

