

Certificate of Analysis


Product MEDONIC M51-L1 LYSE (200 mL)
Product No 1504511
Lot No 2025040201
Manufacturing Date 2025-04-02
Expiration Date 2027-04-01

Test	Specification
pH	2,0 – 2,4
CBC Performance test*	Passes test
Background WBC (x 10 ⁹ /L)	≤ 0,2

*Performance testing at a reference laboratory. WBC and HGB tested and approved.

Actual values meet specification.

Date: 2025-05-28



Pernilla Hammarstedt Rydberg
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

