

Certificate of Analysis

Product BOULE CLEANING KIT (3 x 450 mL)
Product No 1504111
Lot No 2505-319
Manufacturing date 2025-04-22
Use before 2027-04-21


Test Hypochlorite	Specification
Hypochlorite concentration	2.0 – 2.4 %
Blank	PLT < 100 x 10 ⁹ /L RBC < 0.05 x 10 ¹² /L

Test Enzymatic Cleaner	Specification
pH	7.6 – 8.4
Blank	PLT < 100 x 10 ⁹ /L RBC < 0.05 x 10 ¹² /L

Test Detergent Cleaner	Specification
pH	7.6 – 8.4
Blank	PLT < 100 x 10 ⁹ /L RBC < 0.05 x 10 ¹² /L

Actual values meet specifications.

Date: 2025-05-23



Caroline Sparf
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

