

# Certificate of Analysis


**Product** EXIGO LYSE (1.9 L)  
**Product No** 1504202  
**Lot No** 2506-333  
**Manufacturing Date** 2025-06-25  
**Expiration Date** 2028-06-24

Test	Specification
pH	6.95 – 7.25
Osmolality (mOsm/kg)	195 - 225
Conductivity (mS/cm)	12.10 – 12.80
CBC Performance *	Passes test
Particles (PLT/RBC channel)	≤ 5

\* Performance test using calibrated Blood cell counter at Boule Medical reference laboratory. All parameters except mid cells are tested. Accepted reference 95 % to 105 % of expected control blood sample values.

Actual values meet specifications.

Date: 2025-06-27



*Pernilla Hammarstedt Rydberg*  
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.  
All information in this document is property of Boule Medical AB.

